

EC DECLARATION OF CONFORMITY



Name of product: ENDO-TOP

Variant:

ENDO-TOP: 20 pieces, 50 pieces, 100 pieces

Manufacturer:

Przedsiębiorstwo Produkcyjno-Handlowe CERKAMED Wojciech Pawłowski
37-450 Stalowa Wola, ul. Kwiatkowskiego 1, POLAND

Purpose and range of use:

Endo irrigation needles.

Medical device of class IIa, according to the rule 6 of Annex IX MDD 93/42/ EEC.
Evaluation of conformity was conducted following the procedure relating to the EC Declaration of Conformity set out in Annex II excluding p.4 of Directive 93/42/EEC as amended 2007/47/EC.

Reference documents:

- MDD 93/42/ EEC as amended 2007/47/EC
- Annex I of MDD 93/42/EEC as amended 2007/47/EC

Evaluation of conformity was conducted with participation of the notified body:
Országos Gógyszerészeti és Élelmezés-egészségügyi Intézet Eszközminősítő és Kórháztechnikai Igazgatóság. National Institute of Pharmacy and Nutrition (EMKI),
no. ID 1011, Zrínyi utca 3. H-1051, Budapest, Hungary.

Conformity of the product confirmed with certificate No 5-780-200-1601 valid until 2021-01-02.

We declare with full responsibility that the manufactured product, which this statement refers to, complies with the reference documents mentioned above.

Wojciech Pawłowski,
Owner,
Stalowa Wola



.....
signature, company stamp, date

16.04.2018