



# DECLARATION OF CONFORMITY



Manufacturer **W&H Sterilization S.r.l.**

Address Via Bolgara, 2  
I-24060 Brusaporto (BG)  
ITALY

Product **LisaSafe**  
Model Code **A7000150**  
Serial Number **000-001209** (F4010010)  
Technical Specifications Dongle key connected to sterilizer's low voltage power supply;  
Voltage 15 Vcc; Power < 1 W

We herewith declare under our sole responsibility that the above mentioned products meet the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of manufacturer.

Technical file: TF 08


The product is in compliance with the following applicable DIRECTIVES and STANDARDS

#### Directive

- > 2004/108/EC Electromagnetic compatibility (EMC)
- > 2006/95/EC Low Voltage

#### Reference standards:

- > EN 61326 + A1 + A2 + A3 : 1997
- > CISPR 16-2 : 1996
- > CISPR 16-1 : 1993
- > CISPR 22 – EN 55022 + A1 + A2 : 1998
- > EN 61000-4-2 + A1 + A2 : 1995
- > EN 61000-4-3 + A1 + A2 : 1996
- > EN 61000-4-4 : 2004
- > EN 61000-4-5 + A1 : 1995
- > EN 61000-4-6 + A1 : 1996
- > EN 61000-4-8 + A1 : 1993
- > EN 61000-4-11 : 2004

ISSUE DATE	APPROVED BY
31 may 2011	 W&H Sterilization S.r.l. Via Bolgara, 2 - 24060 Brusaporto (BG) Italy Simone Previcini General Management
W&H Sterilization S.r.l. - Via Bolgara, 2 - 24060 Brusaporto (BG) Italy	